

ADAPTOR USE FROM GIVING SETS

1. When you need to flush/medicate during or after feeding

If your tube connects in this way



OR

If your tube connects in this way



2. Disconnect by leaving adaptor connected to tube



OR




3. Connect syringe to clear adaptor and use



OR



4. Disconnect syringe and clear adaptor from tube



OR



5. Now wash and air dry your syringe and adaptor for use next time



CLEAR ADAPTOR CAN BE USED FOR A **24 HOUR** PERIOD